FreeStyle LibreLink

With the FreeStyle Libre 2 system, you can see your glucose readings on your smartphone^{1,2} with a quick glance.



Downloading the FreeStyle LibreLink app¹

- 1 Go to App Store or Play Store
- 2 In the search bar, enter FreeStyle LibreLink



3 Look for FreeStyle LibreLink



4 Press Get or Install
(the app is free to
download and use)³



5 The FreeStyle LibreLink icon will appear on your home screen



How to start the FreeStyle Libre 2 Plus sensor



Open the FreeStyle LibreLink app¹ on your smartphone.

iPhone: Tap the scan button.



iPhone: Touch the sensor with the top of your iPhone.

Android: Hold the back of the Android smartphone near the sensor and position the Near Field Communication (NFC) antenna over the sensor.



iPhone: Do not move your iPhone until you hear a tone and/or feel a vibration.

Android: Continue holding your smartphone near the sensor until you hear a second tone and/or feel a vibration.

This completes the scan.

Once you have set up your FreeStyle Libre 2 Plus sensor your glucose readings are automatically² updated in real-time and sent directly to the FreeStyle LibreLink app¹

QUICK TIPS


- Scan your FreeStyle Libre 2 Plus sensor with your smartphone¹ for glucose readings during a signal loss
- Remove bulky or metallic phone cases as they can interfere with the NFC signal
- A 60-minute warm-up is required when applying the sensor








It's easy to set alarms on your FreeStyle LibreLink app^{1,2}










Tap Alarms in the menu

Touch Low Glucose Alarm and turn on alarm (alarms are off by default)

Signal Loss Alarm is automatically turned on when glucose alarm is turned ON







Alarms

Alarms are only available when you start a Free-Style Libre 2 of Free-Style Libre 2 Plus Sensor with this device.

Low Glucose Alarm On > Below 3.9 mmol/L

High Glucose Alarm Off >

Signal Loss Alarm On >

Scroll to select a Low Glucose Value³

Touch Alarm
Tone and
make your
tone choice

Set your
High Glucose
Alarm using
the same steps⁴

Visit FreeStyle.Abbott for more information

FreeStyle Libre 2



Images are for illustrative purposes only. Not actual patient data.

The FreeStyle LibreLink app is only compatible with certain mobile devices and operating systems. Please check the website for more information about device compatibility before using the app. Use of FreeStyle LibreLink may require registration with LibreView.

1. To receive alarms on your phone, you should start your FreeStyle Libre 2 or FreeStyle Libre 2 Plus sensor with your compatible phone. Notifications will only be received when alarms are turned on and the sensor is within 6 meters unobstructed of the reading device. You must have override do not disturb settings enabled to receive alarms and alerts on your smartphone. 2. Data on file, Abbott Diabetes Care, Inc. 3. 3.9 mmol/L is the default Low Glucose Alarm level and can be set between 3.3-5.6 mmol/L. 4. 13.3 mmol/L is the default High Glucose Alarm level and can be set between 6.7-22.2 mmol/L. © 2024 Abbott. The sensor housing, FreeStyle, Libre, and related brand marks are marks of Abbott. ADC-86944 v1.0 01/24.